

## Quality Control Certificate

Product: **Florisil Column**  
 Product No.: 13807  
 Lot No.: **720208**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The Florisil Column is part of a 3- or 4-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

### Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:	PCDD/F-TEQ:	0,06	pg/column
		(crit: <	0,7 pg/column)
	dl-PCB-TEQ:	0,0253	pg/column
		(crit: <	0,05 pg/column)
	Sum Total PCB:	4,1	pg/column
		(crit: <	300 pg/column)

Results Recoveries:	PCDD/F	70	to	97	%	(crit: 70	to	120	%)
	PCB	86	to	114	%	(crit: 70	to	120	%)

This is to certify that the Florisil Column, Lot 720208, passed the required test specifications and is released for sale.

date: 22.05.2024 sign.: *M. Bradis*

The company LCTech GmbH is certified according to ISO 9001



## QC-Certificate - 13807 - 720208

Hazards:	<p>NOT FOR HUMAN OR DRUG USE!</p> <p>The Florisil Column is designed and prepared for usage with the Universal/standard &amp; Smart Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.</p>
Quality Control:	<p>All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.</p>
Quality Management:	<p>This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)</p>
Documentation / Data Attached:	<p>table 1 &amp; 2: blankvalues of PCDD/F and PCB table 3 &amp; 4: 13C-Recoveries of PCDD/F and PCB</p>
Analytics	<p>This is to certify that the Florisil Column, Lot , passed the required test specifications and is released for sale.</p>
Remarks	<p>Our suppliers maintain the highest standard of quality, however due to the high temperature necessary for several steps in the production, some small charred particles may be visible within a batch of Florisil or filters without any effect on the clean-up.</p>



QC-Certificate - 13807 - 720208

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 11

Table 1: PCDD/F blank

	[pg/column]
2,3,7,8-TCDF	<dl
1,2,3,7,8-PeCDF	<dl
2,3,4,7,8-PeCDF	<dl
1,2,3,4,7,8-HxCDF	<dl
1,2,3,6,7,8-HxCDF	<0,018
2,3,4,6,7,8-HxCDF	<dl
1,2,3,7,8,9-HxCDF	<dl
1,2,3,4,6,7,8-HpCDF	<dl
1,2,3,4,7,8,9-HpCDF	<dl
1,2,3,4,6,7,8,9-OCDF	<dl
2,3,7,8-TCDD	<0,036
1,2,3,7,8-PeCDD	<dl
1,2,3,4,7,8-HxCDD	<dl
1,2,3,6,7,8-HxCDD	<dl
1,2,3,7,8,9-HxCDD	<dl
1,2,3,4,6,7,8-HpCDD	<dl
1,2,3,4,6,7,8,9-OCDD	0,25

Table 2: PCB blank

	[pg/column]
PCB-#28	1,72
PCB-#52	1,4
PCB-#101	0,45
PCB-#153	0,57
PCB-#138	<0,261
PCB-#180	<0,18
PCB-#81	0,17
PCB-#77	0,754
PCB-#126	0,05
PCB-#169	0,671
PCB-#123	0,01
PCB-#118	1,11
PCB-#114	<dl
PCB-#105	0,5
PCB-#167	<dl
PCB-#156	<dl
PCB-#157	0,08
PCB-#189	0,094

PCDD/F TEQ (2005)	[pg/column]
lower bound	0,02
upper bound	0,06

PCB-TEQ	[pg/column]
lower bound	0,0253
upper bound	0,0253
Sum DIN	4,1

Table 3: PCDD/F recoveries

	[%]	RSD [%]	
PCDD/F 13C Recoveries [%]	2,3,7,8-TCDF	91	5
	1,2,3,7,8-PeCDF	89	1
	2,3,4,7,8-PeCDF	92	4
	1,2,3,4,7,8-HxCDF	88	8
	1,2,3,6,7,8-HxCDF	94	9
	2,3,4,6,7,8-HxCDF	93	7
	1,2,3,7,8,9-HxCDF	97	9
	1,2,3,4,6,7,8-HpCDF	79	9
	1,2,3,4,7,8,9-HpCDF	91	8
	1,2,3,4,6,7,8,9-OCDF	77	15
	2,3,7,8-TCDD	82	13
	1,2,3,7,8-PeCDD	89	8
	1,2,3,4,7,8-HxCDD	94	9
	1,2,3,6,7,8-HxCDD	82	8
	1,2,3,7,8,9-HxCDD	94	12
	1,2,3,4,6,7,8-HpCDD	81	9
	1,2,3,4,6,7,8,9-OCDD	70	11

Table 4: PCB recoveries

	[%]	RSD [%]	
PCB 13C Recoveries [%]	PCB-#28	95	2
	PCB-#52	86	6
	PCB-#101	93	5
	PCB-#153	97	4
	PCB-#138	99	3
	PCB-#180	101	3
	PCB-#81	101	8
	PCB-#77	106	9
	PCB-#126	111	11
	PCB-#169	105	16
	PCB-#123	114	9
	PCB-#118	98	6
	PCB-#114	100	8
	PCB-#105	101	8
	PCB-#167	104	3
	PCB-#156	102	9
	PCB-#157	99	8
	PCB-#189	107	5