

Quality Control Certificate

Product: **Alumina Column**
 Product No.: 15433
 Lot No.: **719889**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The Alumina Column is part of a 3-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.


Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:	PCDD/F-TEQ:	0,05	pg/column
		(crit: <	0,7 pg/column)
	dl-PCB-TEQ:	0	pg/column
		(crit: <	0,05 pg/column)
	Sum Total PCB:	3,4	pg/column
		(crit: <	300 pg/column)

Results Recoveries:	PCDD/F	96	to	116	%	(crit: 70	to	120	%)
	PCB	95	to	106	%	(crit: 70	to	120	%)

This is to certify that the Alumina Column, Lot 719889, passed the required test specifications and is released for sale.

date: 03.04.2024 sign.: 

The company LCTech GmbH is certified according to ISO 9001



QC-Certificate - 15433 - 719889

Hazards:	<p>NOT FOR HUMAN OR DRUG USE!</p> <p>The Alumina Column is designed and prepared for usage with the Universal/standard & Smart Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.</p>
Quality Control:	<p>All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.</p>
Quality Management:	<p>This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)</p>
Documentation / Data Attached:	<p>table 1 & 2: blankvalues of PCDD/F and PCB table 3 & 4: 13C-Recoveries of PCDD/F and PCB</p>
Analytics	<p>This is to certify that the Alumina Column, Lot , passed the required test specifications and is released for sale.</p>
Remarks	<p>n/a</p>



QC-Certificate - 15433 - 719889

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 6

Table 1: PCDD/F blank

	[pg/column]
2,3,7,8-TCDF	<dl
1,2,3,7,8-PeCDF	<dl
2,3,4,7,8-PeCDF	<dl
1,2,3,4,7,8-HxCDF	<dl
1,2,3,6,7,8-HxCDF	<dl
2,3,4,6,7,8-HxCDF	<dl
1,2,3,7,8,9-HxCDF	<dl
1,2,3,4,6,7,8-HpCDF	<dl
1,2,3,4,7,8,9-HpCDF	<dl
1,2,3,4,6,7,8,9-OCDF	<dl
2,3,7,8-TCDD	<dl
1,2,3,7,8-PeCDD	<dl
1,2,3,4,7,8-HxCDD	<dl
1,2,3,6,7,8-HxCDD	<dl
1,2,3,7,8,9-HxCDD	<dl
1,2,3,4,6,7,8-HpCDD	<dl
1,2,3,4,6,7,8,9-OCDD	0,15

Table 2: PCB blank

	[pg/column]
PCB-#28	1,27
PCB-#52	1,57
PCB-#101	0,51
PCB-#153	<dl
PCB-#138	<dl
PCB-#180	<dl
PCB-#81	<dl
PCB-#77	<dl
PCB-#126	<dl
PCB-#169	<dl
PCB-#123	<dl
PCB-#118	<dl
PCB-#114	<dl
PCB-#105	<dl
PCB-#167	<dl
PCB-#156	<dl
PCB-#157	<dl
PCB-#189	<dl

PCDD/F TEQ (2005)	[pg/column]
lower bound	0
upper bound	0,05

PCB-TEQ	[pg/column]
lower bound	0
upper bound	0,0004
Sum DIN	3,4

Table 3: PCDD/F recoveries

	[%]	RSD [%]
2,3,7,8-TCDF	101	4
1,2,3,7,8-PeCDF	99	4
2,3,4,7,8-PeCDF	105	3
1,2,3,4,7,8-HxCDF	102	4
1,2,3,6,7,8-HxCDF	113	4
2,3,4,6,7,8-HxCDF	116	1
1,2,3,7,8,9-HxCDF	111	3
1,2,3,4,6,7,8-HpCDF	110	4
1,2,3,4,7,8,9-HpCDF	97	4
1,2,3,4,6,7,8,9-OCDF	107	4
2,3,7,8-TCDD	96	6
1,2,3,7,8-PeCDD	102	6
1,2,3,4,7,8-HxCDD	111	3
1,2,3,6,7,8-HxCDD	97	4
1,2,3,7,8,9-HxCDD	113	3
1,2,3,4,6,7,8-HpCDD	101	3
1,2,3,4,6,7,8,9-OCDD	103	4

Table 4: PCB recoveries

	[%]	RSD [%]
PCB-#28	104	3
PCB-#52	101	8
PCB-#101	95	3
PCB-#153	102	6
PCB-#138	99	3
PCB-#180	105	2
PCB-#81	100	4
PCB-#77	104	4
PCB-#126	105	7
PCB-#169	101	6
PCB-#123	96	2
PCB-#118	101	3
PCB-#114	97	4
PCB-#105	95	7
PCB-#167	103	1
PCB-#156	98	3
PCB-#157	104	3
PCB-#189	106	4