

## Quality Control Certificate

Product: **Alumina Column**  
 Product No.: 15433  
 Lot No.: **720367**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The Alumina Column is part of a 3-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

### Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:	PCDD/F-TEQ:	0,05	pg/column
		(crit: <	0,7 pg/column)
	dl-PCB-TEQ:	0,0329	pg/column
		(crit: <	0,05 pg/column)
	Sum Total PCB:	3,2	pg/column
		(crit: <	300 pg/column)

Results Recoveries:	PCDD/F	71	to	115	%	(crit: 70	to	120	%)
	PCB	89	to	102	%	(crit: 70	to	120	%)

This is to certify that the Alumina Column, Lot 720367, passed the required test specifications and is released for sale.

date: 19.06.2024 sign.: *M. Brack*

The company LCTech GmbH is certified according to ISO 9001



## QC-Certificate - 15433 - 720367

Hazards:	<p>NOT FOR HUMAN OR DRUG USE!</p> <p>The Alumina Column is designed and prepared for usage with the Universal/standard &amp; Smart Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.</p>
Quality Control:	<p>All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.</p>
Quality Management:	<p>This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)</p>
Documentation / Data Attached:	<p>table 1 &amp; 2: blankvalues of PCDD/F and PCB table 3 &amp; 4: 13C-Recoveries of PCDD/F and PCB</p>
Analytics	<p>This is to certify that the Alumina Column, Lot , passed the required test specifications and is released for sale.</p>
Remarks	<p>n/a</p>



QC-Certificate - 15433 - 720367

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 9

Table 1: PCDD/F blank

	[pg/column]
2,3,7,8-TCDF	<dl
1,2,3,7,8-PeCDF	<dl
2,3,4,7,8-PeCDF	<dl
1,2,3,4,7,8-HxCDF	<dl
1,2,3,6,7,8-HxCDF	<dl
2,3,4,6,7,8-HxCDF	<dl
1,2,3,7,8,9-HxCDF	<dl
1,2,3,4,6,7,8-HpCDF	<dl
1,2,3,4,7,8,9-HpCDF	<dl
1,2,3,4,6,7,8,9-OCDF	<0,054
2,3,7,8-TCDD	<dl
1,2,3,7,8-PeCDD	<dl
1,2,3,4,7,8-HxCDD	<0,027
1,2,3,6,7,8-HxCDD	<0,108
1,2,3,7,8,9-HxCDD	<dl
1,2,3,4,6,7,8-HpCDD	<0,09
1,2,3,4,6,7,8,9-OCDD	0,41

Table 2: PCB blank

	[pg/column]
PCB-#28	1,57
PCB-#52	0,46
PCB-#101	0,19
PCB-#153	<0,162
PCB-#138	0,59
PCB-#180	0,4
PCB-#81	0,59
PCB-#77	0,7433
PCB-#126	0,17
PCB-#169	0,52
PCB-#123	0,11
PCB-#118	0,15
PCB-#114	0,062
PCB-#105	<0,081
PCB-#167	<dl
PCB-#156	0,191
PCB-#157	0,12
PCB-#189	<0,0072

PCDD/F TEQ (2005)	[pg/column]
lower bound	0,01
upper bound	0,05

PCB-TEQ	[pg/column]
lower bound	0,0329
upper bound	0,0329
Sum DIN	3,2

Table 3: PCDD/F recoveries

	[%]	RSD [%]	
PCDD/F 13C Recoveries [%]	2,3,7,8-TCDF	77	12
	1,2,3,7,8-PeCDF	73	11
	2,3,4,7,8-PeCDF	76	10
	1,2,3,4,7,8-HxCDF	94	9
	1,2,3,6,7,8-HxCDF	95	9
	2,3,4,6,7,8-HxCDF	106	6
	1,2,3,7,8,9-HxCDF	110	4
	1,2,3,4,6,7,8-HpCDF	109	8
	1,2,3,4,7,8,9-HpCDF	101	10
	1,2,3,4,6,7,8,9-OCDF	99	8
	2,3,7,8-TCDD	71	16
	1,2,3,7,8-PeCDD	78	11
	1,2,3,4,7,8-HxCDD	111	9
	1,2,3,6,7,8-HxCDD	90	8
	1,2,3,7,8,9-HxCDD	115	7
	1,2,3,4,6,7,8-HpCDD	104	9
	1,2,3,4,6,7,8,9-OCDD	101	6

Table 4: PCB recoveries

	[%]	RSD [%]	
PCB 13C Recoveries [%]	PCB-#28	101	6
	PCB-#52	98	10
	PCB-#101	94	6
	PCB-#153	95	5
	PCB-#138	96	6
	PCB-#180	95	3
	PCB-#81	93	5
	PCB-#77	91	7
	PCB-#126	98	10
	PCB-#169	89	7
	PCB-#123	102	4
	PCB-#118	100	4
	PCB-#114	97	5
	PCB-#105	95	8
	PCB-#167	93	2
	PCB-#156	97	5
	PCB-#157	98	4
	PCB-#189	93	4