

Quality Control Certificate

Product: **Alumina Column**
Product No.: 15433
Lot No.: **721001**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The Alumina Column is part of a 3-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of $R > 10000$.

Results Blank Value:

PCDD/F-TEQ:	0,08	pg/column
	(crit: <	0,70 pg/column)
dl-PCB-TEQ:	0,001	pg/column
	(crit: <	0,05 pg/column)
Sum Total PCB:	0	pg/column
	(crit: <	300 pg/column)

Results Recoveries:

PCDD/F	86	to	114	%	(crit: 70	to	120	%)
PCB	95	to	106	%	(crit: 70	to	120	%)

This is to certify that the Alumina Column, Lot 721001, passed the required test specifications and is released for sale.

date: 14.11.2024

sign.: 

Michael Brandis

The company LCTech GmbH is certified according to ISO 9001



QC-Certificate - 15433 - 721001

Hazards:	NOT FOR HUMAN OR DRUG USE! The Alumina Column is designed and prepared for usage with the Universal/standard & Smart Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.
Quality Control:	All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.
Quality Management:	This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)
Documentation / Data Attached:	table 1 & 2: blankvalues of PCDD/F and PCB table 3 & 4: 13C-Recoveries of PCDD/F and PCB
Analytics	This is to certify that the Alumina Column, Lot , passed the required test specifications and is released for sale.
Remarks	n/a

QC-Certificate - 15433 - 721001

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 6

Table 1: PCDD/F blank

	[pg/column]
2,3,7,8-TCDF	<dl
1,2,3,7,8-PeCDF	<dl
2,3,4,7,8-PeCDF	<dl
1,2,3,4,7,8-HxCDF	<dl
1,2,3,6,7,8-HxCDF	0,027
2,3,4,6,7,8-HxCDF	<0,045
1,2,3,7,8,9-HxCDF	0,08
1,2,3,4,6,7,8-HpCDF	<0,063
1,2,3,4,7,8,9-HpCDF	0,071
1,2,3,4,6,7,8,9-OCDF	<0,054
2,3,7,8-TCDD	<dl
1,2,3,7,8-PeCDD	<dl
1,2,3,4,7,8-HxCDD	0,047
1,2,3,6,7,8-HxCDD	<0,108
1,2,3,7,8,9-HxCDD	0,148
1,2,3,4,6,7,8-HpCDD	<0,09
1,2,3,4,6,7,8,9-OCDD	0,73

Table 2: PCB blank

	[pg/column]
PCB-#28	<dl
PCB-#52	<dl
PCB-#101	<dl
PCB-#153	<dl
PCB-#138	<dl
PCB-#180	<dl
PCB-#81	<dl
PCB-#77	<dl
PCB-#126	0,01
PCB-#169	<dl
PCB-#123	<dl
PCB-#118	<dl
PCB-#114	<dl
PCB-#105	<dl
PCB-#167	<dl
PCB-#156	<dl
PCB-#157	<0,018
PCB-#189	<dl

PCDD/F TEQ (2005)	[pg/column]
lower bound	0,04
upper bound	0,08

PCB-TEQ	[pg/column]
lower bound	0,001
upper bound	0,0013
Sum DIN	0

Table 3: PCDD/F recoveries

	[%]	RSD [%]	
PCDD/F 13C Recoveries [%]	2,3,7,8-TCDF	96	5
	1,2,3,7,8-PeCDF	100	4
	2,3,4,7,8-PeCDF	103	9
	1,2,3,4,7,8-HxCDF	86	13
	1,2,3,6,7,8-HxCDF	95	13
	2,3,4,6,7,8-HxCDF	92	13
	1,2,3,7,8,9-HxCDF	98	11
	1,2,3,4,6,7,8-HpCDF	105	3
	1,2,3,4,7,8,9-HpCDF	114	7
	1,2,3,4,6,7,8,9-OCDF	96	13
	2,3,7,8-TCDD	91	6
	1,2,3,7,8-PeCDD	106	5
	1,2,3,4,7,8-HxCDD	99	13
	1,2,3,6,7,8-HxCDD	88	17
	1,2,3,7,8,9-HxCDD	96	12
	1,2,3,4,6,7,8-HpCDD	106	3
1,2,3,4,6,7,8,9-OCDD	103	8	

Table 4: PCB recoveries

	[%]	RSD [%]	
PCB 13C Recoveries [%]	PCB-#28	104	3
	PCB-#52	98	9
	PCB-#101	96	4
	PCB-#153	100	2
	PCB-#138	98	5
	PCB-#180	102	5
	PCB-#81	97	2
	PCB-#77	96	2
	PCB-#126	95	8
	PCB-#169	97	9
	PCB-#123	99	4
	PCB-#118	102	4
	PCB-#114	96	5
	PCB-#105	97	6
	PCB-#167	106	10
	PCB-#156	100	10
	PCB-#157	105	11
	PCB-#189	105	4