

## Quality Control Certificate

Product: **Alumina Column**  
 Product No.: 15433  
 Lot No.: **718777**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The Alumina Column is part of a 3-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

### Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:	PCDD/F-TEQ:	0,1	pg/column
		(crit: <	0,7 pg/column)
	dl-PCB-TEQ:	0,0144	pg/column
	(crit: <	0,05 pg/column)	
	Sum Total PCB:	2,7	pg/column
	(crit: <	300 pg/column)	

Results Recoveries:	PCDD/F	90	to	117	%	(crit: 70	to	120	%)
	PCB	99	to	116	%	(crit: 70	to	120	%)

This is to certify that the Alumina Column, Lot 718777, passed the required test specifications and is released for sale.

date: 26.09.2023 sign.: *M. Bradis*

The company LCTech GmbH is certified according to ISO 9001



## QC-Certificate - 15433 - 718777

Hazards:	<p>NOT FOR HUMAN OR DRUG USE!</p> <p>The Alumina Column is designed and prepared for usage with the Universal/standard &amp; Smart Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.</p>
Quality Control:	<p>All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.</p>
Quality Management:	<p>This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)</p>
Documentation / Data Attached:	<p>table 1 &amp; 2: blankvalues of PCDD/F and PCB table 3 &amp; 4: 13C-Recoveries of PCDD/F and PCB</p>
Analytics	<p>This is to certify that the Alumina Column, Lot , passed the required test specifications and is released for sale.</p>
Remarks	<p>n/a</p>



QC-Certificate - 15433 - 718777

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 6

Table 1: PCDD/F blank

	[pg/column]
2,3,7,8-TCDF	<dl
1,2,3,7,8-PeCDF	<dl
2,3,4,7,8-PeCDF	<0,081
1,2,3,4,7,8-HxCDF	0,033
1,2,3,6,7,8-HxCDF	0,032
2,3,4,6,7,8-HxCDF	<0,045
1,2,3,7,8,9-HxCDF	<0,045
1,2,3,4,6,7,8-HpCDF	0,1
1,2,3,4,7,8,9-HpCDF	0,101
1,2,3,4,6,7,8,9-OCDF	0,11
2,3,7,8-TCDD	<0,036
1,2,3,7,8-PeCDD	<dl
1,2,3,4,7,8-HxCDD	0,051
1,2,3,6,7,8-HxCDD	0,13
1,2,3,7,8,9-HxCDD	0,033
1,2,3,4,6,7,8-HpCDD	0,15
1,2,3,4,6,7,8,9-OCDD	0,7

Table 2: PCB blank

	[pg/column]
PCB-#28	0,79
PCB-#52	0,54
PCB-#101	0,49
PCB-#153	0,26
PCB-#138	0,17
PCB-#180	0,453
PCB-#81	0,22
PCB-#77	0,2417
PCB-#126	0,07
PCB-#169	0,242
PCB-#123	<0,18
PCB-#118	0,3
PCB-#114	<0,045
PCB-#105	0,1
PCB-#167	0,096
PCB-#156	0,312
PCB-#157	0,08
PCB-#189	0,468

PCDD/F TEQ (2005)	[pg/column]
lower bound	0,07
upper bound	0,1

PCB-TEQ	[pg/column]
lower bound	0,0144
upper bound	0,0144
Sum DIN	2,7

Table 3: PCDD/F recoveries

	[%]	RSD [%]	
PCDD/F 13C Recoveries [%]	2,3,7,8-TCDF	101	3
	1,2,3,7,8-PeCDF	90	9
	2,3,4,7,8-PeCDF	93	16
	1,2,3,4,7,8-HxCDF	114	7
	1,2,3,6,7,8-HxCDF	117	3
	2,3,4,6,7,8-HxCDF	112	7
	1,2,3,7,8,9-HxCDF	109	4
	1,2,3,4,6,7,8-HpCDF	113	4
	1,2,3,4,7,8,9-HpCDF	106	9
	1,2,3,4,6,7,8,9-OCDF	105	6
	2,3,7,8-TCDD	95	8
	1,2,3,7,8-PeCDD	92	8
	1,2,3,4,7,8-HxCDD	113	4
	1,2,3,6,7,8-HxCDD	97	4
	1,2,3,7,8,9-HxCDD	114	3
	1,2,3,4,6,7,8-HpCDD	105	4
	1,2,3,4,6,7,8,9-OCDD	99	3

Table 4: PCB recoveries

	[%]	RSD [%]	
PCB 13C Recoveries [%]	PCB-#28	99	5
	PCB-#52	101	6
	PCB-#101	103	2
	PCB-#153	100	4
	PCB-#138	101	3
	PCB-#180	112	4
	PCB-#81	116	0
	PCB-#77	116	0
	PCB-#126	114	0
	PCB-#169	116	0
	PCB-#123	104	4
	PCB-#118	104	3
	PCB-#114	104	4
	PCB-#105	105	5
	PCB-#167	105	1
	PCB-#156	108	2
	PCB-#157	112	2
	PCB-#189	111	2