

Quality Control Certificate

Product: **Universal Column**
Product No.: 19511
Lot No.: **718172**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The Universal Column is part of a 3- or 4-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:

PCDD/F-TEQ:	0,2	pg/column
	(crit: <	0,7 pg/column)
dl-PCB-TEQ:	0,0115	pg/column
	(crit: <	0,05 pg/column)
Sum Total PCB:	36,4	pg/column
	(crit: <	300 pg/column)

Results Recoveries:

PCDD/F	88	to	117	%	(crit: 70	to	120	%)
PCB	71	to	107	%	(crit: 70	to	120	%)

This is to certify that the Universal Column, Lot 718172, passed the required test specifications and is released for sale.

date: 27.03.2023 sign.: T. Keshmeir

The company LCTech GmbH is certified according to ISO 9001



QC-Certificate - 19511 - 718172

Hazards:	<p>NOT FOR HUMAN OR DRUG USE!</p> <p>The 209 Column is designed and prepared for usage with the Alumina Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.</p>
Quality Control:	<p>All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.</p>
Quality Management:	<p>This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)</p>
Documentation / Data Attached:	<p>table 1 & 2: blankvalues of PCDD/F and PCB table 3 & 4: 13C-Recoveries of PCDD/F and PCB</p>
Analytics	<p>All the Columns (n>5) have to perform a clean-up of a solvent blank (10 mL n-hexane), spiked with a 13C - labelled quantifier-standard solution with a default alumina plus or pure 209 method onto a DEXTech Pure or Plus system. There are 2 fractions, fraction 1 (all 209 PCB) and fraction 2 (PCDD/F). Both fractions are spiked with the corresponding 13C - labelled recovery- standard solutions and evaporated with the D-EVA vacuum centrifuge. The extracts are measured with a HRMS-DFS from Thermo Fisher Scientific with a resolution of R > 10000. The HRGCs are equipped with 60 m DB5 MS Columns. For PCDD/F 5µL are injected via PTV, for PCB</p>
Remarks	<p>Our suppliers maintain the highest standard of quality, however due to the high temperature necessary for several steps in the production, some small charred particles may be visible within a batch of silica or filters without any effect on the clean-up.</p>

QC-Certificate - 19511 - 718172

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 4

Table 1: PCDD/F blank

	[pg/column]
2,3,7,8-TCDF	0,08
1,2,3,7,8-PeCDF	0,07
2,3,4,7,8-PeCDF	<0,081
1,2,3,4,7,8-HxCDF	0,052
1,2,3,6,7,8-HxCDF	0,02
2,3,4,6,7,8-HxCDF	<0,045
1,2,3,7,8,9-HxCDF	<0,045
1,2,3,4,6,7,8-HpCDF	0,23
1,2,3,4,7,8,9-HpCDF	0,032
1,2,3,4,6,7,8,9-OCDF	0,25
2,3,7,8-TCDD	<dl
1,2,3,7,8-PeCDD	0,11
1,2,3,4,7,8-HxCDD	0,03
1,2,3,6,7,8-HxCDD	0,2
1,2,3,7,8,9-HxCDD	0,038
1,2,3,4,6,7,8-HpCDD	0,3
1,2,3,4,6,7,8,9-OCDD	3

Table 2: PCB blank

	[pg/column]
PCB-#28	6,61
PCB-#52	8,02
PCB-#101	7,17
PCB-#153	6,48
PCB-#138	5,89
PCB-#180	2,275
PCB-#81	0,09
PCB-#77	0,268
PCB-#126	0,0758
PCB-#169	0,121
PCB-#123	0,41
PCB-#118	2,91
PCB-#114	0,671
PCB-#105	1,34
PCB-#167	0,762
PCB-#156	0,703
PCB-#157	0,55
PCB-#189	1,352

PCDD/F TEQ (2005)	[pg/column]
lower bound	0,19
upper bound	0,2

PCB-TEQ	[pg/column]
lower bound	0,0115
upper bound	0,0115
Sum DIN	36,4

QC-Certificate - 19511 - 718172

Table 3: PCDD/F recoveries

	[%]	RSD [%]	
PCDD/F 13C Recoveries [%]	2,3,7,8-TCDF	107	4
	1,2,3,7,8-PeCDF	105	4
	2,3,4,7,8-PeCDF	105	5
	1,2,3,4,7,8-HxCDF	98	6
	1,2,3,6,7,8-HxCDF	104	7
	2,3,4,6,7,8-HxCDF	98	7
	1,2,3,7,8,9-HxCDF	112	5
	1,2,3,4,6,7,8-HpCDF	116	5
	1,2,3,4,7,8,9-HpCDF	117	0
	1,2,3,4,6,7,8,9-OCDF	110	9
	2,3,7,8-TCDD	98	4
	1,2,3,7,8-PeCDD	99	6
	1,2,3,4,7,8-HxCDD	106	5
	1,2,3,6,7,8-HxCDD	88	4
	1,2,3,7,8,9-HxCDD	97	6
	1,2,3,4,6,7,8-HpCDD	115	9
	1,2,3,4,6,7,8,9-OCDD	98	8

Table 4: PCB recoveries

	[%]	RSD [%]	
PCB 13C Recoveries [%]	PCB-#28	90	7
	PCB-#52	92	7
	PCB-#101	97	3
	PCB-#153	93	3
	PCB-#138	97	5
	PCB-#180	92	2
	PCB-#81	97	3
	PCB-#77	107	6
	PCB-#126	106	5
	PCB-#169	105	6
	PCB-#123	82	6
	PCB-#118	75	9
	PCB-#114	91	3
	PCB-#105	84	6
	PCB-#167	71	10
	PCB-#156	79	6
	PCB-#157	76	8
	PCB-#189	76	5