

Quality Control Certificate

Product: **Universal Column**
 Product No.: 19511
 Lot No.: **721268**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The Universal Column is part of a 3- or 4-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.


Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:	PCDD/F-TEQ:	0,06	pg/column
		(crit: <	0,70 pg/column)
	dl-PCB-TEQ:	0,0029	pg/column
		(crit: <	0,05 pg/column)
	Sum Total PCB:	2,1	pg/column
		(crit: <	300 pg/column)

Results Recoveries:	PCDD/F	77	to	106	%	(crit: 70	to	120	%)
	PCB	73	to	107	%	(crit: 70	to	120	%)

This is to certify that the Universal Column, Lot 721268, passed the required test specifications and is released for sale.

date: 29.01.2025 sign.: 

The company LCTech GmbH is certified according to ISO 9001



QC-Certificate - 19511 - 721268

Hazards:	<p>NOT FOR HUMAN OR DRUG USE!</p> <p>The Universal Column is designed and prepared for usage with the Alumina/Florisol Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.</p>
Quality Control:	<p>All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.</p>
Quality Management:	<p>This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)</p>
Documentation / Data Attached:	<p>table 1 & 2: blankvalues of PCDD/F and PCB table 3 & 4: 13C-Recoveries of PCDD/F and PCB</p>
Analytics	<p>This is to certify that the Universal Column, Lot , passed the required test specifications and is released for sale.</p>
Remarks	<p>Our suppliers maintain the highest standard of quality, however due to the high temperature necessary for several steps in the production, some small charred particles may be visible within a batch of silica or filters without any effect on the clean-up.</p>



QC-Certificate - 19511 - 721268

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 6

Table 1: PCDD/F blank

	[pg/column]
2,3,7,8-TCDF	<0,036
1,2,3,7,8-PeCDF	<dl
2,3,4,7,8-PeCDF	<dl
1,2,3,4,7,8-HxCDF	<0,027
1,2,3,6,7,8-HxCDF	<0,018
2,3,4,6,7,8-HxCDF	<0,045
1,2,3,7,8,9-HxCDF	<dl
1,2,3,4,6,7,8-HpCDF	<dl
1,2,3,4,7,8,9-HpCDF	<dl
1,2,3,4,6,7,8,9-OCDF	<0,054
2,3,7,8-TCDD	<dl
1,2,3,7,8-PeCDD	<dl
1,2,3,4,7,8-HxCDD	<dl
1,2,3,6,7,8-HxCDD	<dl
1,2,3,7,8,9-HxCDD	<dl
1,2,3,4,6,7,8-HpCDD	0,09
1,2,3,4,6,7,8,9-OCDD	0,49

Table 2: PCB blank

	[pg/column]
PCB-#28	0,67
PCB-#52	1,1
PCB-#101	<0,18
PCB-#153	0,37
PCB-#138	<0,261
PCB-#180	<dl
PCB-#81	0,03
PCB-#77	0,078
PCB-#126	0,0292
PCB-#169	<dl
PCB-#123	0,01
PCB-#118	0,22
PCB-#114	0,002
PCB-#105	0,09
PCB-#167	<dl
PCB-#156	<dl
PCB-#157	0,05
PCB-#189	0,027

PCDD/F TEQ (2005)	[pg/column]
lower bound	0,01
upper bound	0,06

PCB-TEQ	[pg/column]
lower bound	0,0029
upper bound	0,0033
Sum DIN	2,1



QC-Certificate - 19511 - 721268

Table 3: PCDD/F recoveries

	[%]	RSD [%]	
PCDD/F 13C Recoveries [%]	2,3,7,8-TCDF	95	4
	1,2,3,7,8-PeCDF	96	8
	2,3,4,7,8-PeCDF	102	8
	1,2,3,4,7,8-HxCDF	84	12
	1,2,3,6,7,8-HxCDF	91	7
	2,3,4,6,7,8-HxCDF	91	9
	1,2,3,7,8,9-HxCDF	97	7
	1,2,3,4,6,7,8-HpCDF	97	8
	1,2,3,4,7,8,9-HpCDF	106	6
	1,2,3,4,6,7,8,9-OCDF	106	7
	2,3,7,8-TCDD	95	4
	1,2,3,7,8-PeCDD	106	9
	1,2,3,4,7,8-HxCDD	87	13
	1,2,3,6,7,8-HxCDD	77	13
	1,2,3,7,8,9-HxCDD	96	10
	1,2,3,4,6,7,8-HpCDD	93	9
1,2,3,4,6,7,8,9-OCDD	88	9	

Table 4: PCB recoveries

	[%]	RSD [%]	
PCB 13C Recoveries [%]	PCB-#28	87	6
	PCB-#52	73	12
	PCB-#101	92	2
	PCB-#153	98	7
	PCB-#138	107	4
	PCB-#180	93	6
	PCB-#81	95	5
	PCB-#77	99	4
	PCB-#126	103	6
	PCB-#169	106	5
	PCB-#123	97	11
	PCB-#118	90	11
	PCB-#114	96	7
	PCB-#105	87	12
	PCB-#167	78	10
	PCB-#156	81	10
	PCB-#157	77	14
	PCB-#189	79	10