

Quality Control Certificate

Product: **Smart Column**
 Product No.: 19513
 Lot No.: **721269**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The Smart Column is part of a 3-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:	PCDD/F-TEQ:	0,05	pg/column
		(crit: <	0,70 pg/column)
	dl-PCB-TEQ:	0,0046	pg/column
		(crit: <	0,05 pg/column)
	Sum Total PCB:	6	pg/column
		(crit: <	300 pg/column)

Results Recoveries:	PCDD/F	86	to	103	%	(crit: 70	to	120	%)
	PCB	72	to	110	%	(crit: 70	to	120	%)

This is to certify that the Smart Column, Lot 721269, passed the required test specifications and is released for sale.

date: 08.01.2025 sign.: *M. Brack's*

The company LCTech GmbH is certified according to ISO 9001



QC-Certificate - 19513 - 721269

Hazards:	<p>NOT FOR HUMAN OR DRUG USE!</p> <p>The Smart Column is designed and prepared for usage with the Alumina/Florisil Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.</p>
Quality Control:	<p>All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.</p>
Quality Management:	<p>This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)</p>
Documentation / Data Attached:	<p>table 1 & 2: blankvalues of PCDD/F and PCB table 3 & 4: 13C-Recoveries of PCDD/F and PCB</p>
Analytics	<p>This is to certify that the Smart Column, Lot , passed the required test specifications and is released for sale.</p>
Remarks	<p>Our suppliers maintain the highest standard of quality, however due to the high temperature necessary for several steps in the production, some small charred particles may be visible within a batch of silica or filters without any effect on the clean-up.</p>



QC-Certificate - 19513 - 721269

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 6

Table 1: PCDD/F blank

	[pg/column]
2,3,7,8-TCDF	<0,036
1,2,3,7,8-PeCDF	<dl
2,3,4,7,8-PeCDF	<dl
1,2,3,4,7,8-HxCDF	<dl
1,2,3,6,7,8-HxCDF	<dl
2,3,4,6,7,8-HxCDF	<dl
1,2,3,7,8,9-HxCDF	<dl
1,2,3,4,6,7,8-HpCDF	<dl
1,2,3,4,7,8,9-HpCDF	<dl
1,2,3,4,6,7,8,9-OCDF	<0,054
2,3,7,8-TCDD	<dl
1,2,3,7,8-PeCDD	<0,054
1,2,3,4,7,8-HxCDD	<0,027
1,2,3,6,7,8-HxCDD	<dl
1,2,3,7,8,9-HxCDD	<0,027
1,2,3,4,6,7,8-HpCDD	<0,09
1,2,3,4,6,7,8,9-OCDD	0,38

Table 2: PCB blank

	[pg/column]
PCB-#28	0,52
PCB-#52	0,96
PCB-#101	2,64
PCB-#153	1,29
PCB-#138	0,28
PCB-#180	0,352
PCB-#81	<0,027
PCB-#77	<dl
PCB-#126	0,0456
PCB-#169	<dl
PCB-#123	<dl
PCB-#118	0,26
PCB-#114	0,034
PCB-#105	<dl
PCB-#167	<dl
PCB-#156	<dl
PCB-#157	0,08
PCB-#189	0,151

PCDD/F TEQ (2005)	[pg/column]
lower bound	0,03
upper bound	0,05

PCB-TEQ	[pg/column]
lower bound	0,0046
upper bound	0,0049
Sum DIN	6

Table 3: PCDD/F recoveries

	[%]	RSD [%]	
PCDD/F 13C Recoveries [%]	2,3,7,8-TCDF	90	3
	1,2,3,7,8-PeCDF	92	5
	2,3,4,7,8-PeCDF	95	4
	1,2,3,4,7,8-HxCDF	93	4
	1,2,3,6,7,8-HxCDF	101	4
	2,3,4,6,7,8-HxCDF	97	4
	1,2,3,7,8,9-HxCDF	100	5
	1,2,3,4,6,7,8-HpCDF	101	5
	1,2,3,4,7,8,9-HpCDF	103	4
	1,2,3,4,6,7,8,9-OCDF	103	6
	2,3,7,8-TCDD	95	3
	1,2,3,7,8-PeCDD	89	5
	1,2,3,4,7,8-HxCDD	100	5
	1,2,3,6,7,8-HxCDD	86	4
	1,2,3,7,8,9-HxCDD	103	4
	1,2,3,4,6,7,8-HpCDD	93	4
	1,2,3,4,6,7,8,9-OCDD	97	6

Table 4: PCB recoveries

	[%]	RSD [%]	
PCB 13C Recoveries [%]	PCB-#28	92	2
	PCB-#52	110	2
	PCB-#101	102	3
	PCB-#153	99	4
	PCB-#138	98	3
	PCB-#180	90	2
	PCB-#81	94	2
	PCB-#77	97	3
	PCB-#126	100	4
	PCB-#169	98	5
	PCB-#123	81	7
	PCB-#118	80	8
	PCB-#114	80	4
	PCB-#105	82	7
	PCB-#167	79	9
	PCB-#156	76	7
	PCB-#157	72	8
	PCB-#189	82	5