

## Quality Control Certificate

Product: **Smart Column**  
 Product No.: 19513  
 Lot No.: **721718**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The Smart Column is part of a 3-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.


### Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:	PCDD/F-TEQ:	0,08	pg/column
		(crit: <	0,70 pg/column)
	dl-PCB-TEQ:	0,002	pg/column
		(crit: <	0,05 pg/column)
	Sum Total PCB:	2,1	pg/column
		(crit: <	300 pg/column)

Results Recoveries:	PCDD/F	85	to	105	%	(crit: 70	to	120	%)
	PCB	74	to	99	%	(crit: 70	to	120	%)

This is to certify that the Smart Column, Lot 721718, passed the required test specifications and is released for sale.

date: 31.03.2025 sign.: 

The company LCTech GmbH is certified according to ISO 9001



## QC-Certificate - 19513 - 721718

Hazards:	<p>NOT FOR HUMAN OR DRUG USE!</p> <p>The Smart Column is designed and prepared for usage with the Alumina/Florisil Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.</p>
Quality Control:	<p>All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.</p>
Quality Management:	<p>This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)</p>
Documentation / Data Attached:	<p>table 1 &amp; 2: blankvalues of PCDD/F and PCB table 3 &amp; 4: 13C-Recoveries of PCDD/F and PCB</p>
Analytics	<p>This is to certify that the Smart Column, Lot , passed the required test specifications and is released for sale.</p>
Remarks	<p>Our suppliers maintain the highest standard of quality, however due to the high temperature necessary for several steps in the production, some small charred particles may be visible within a batch of silica or filters without any effect on the clean-up.</p>

QC-Certificate - 19513 - 721718

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 6

Table 1: PCDD/F blank

	[pg/column]
2,3,7,8-TCDF	<dl
1,2,3,7,8-PeCDF	<0,045
2,3,4,7,8-PeCDF	<dl
1,2,3,4,7,8-HxCDF	<dl
1,2,3,6,7,8-HxCDF	<dl
2,3,4,6,7,8-HxCDF	<dl
1,2,3,7,8,9-HxCDF	<dl
1,2,3,4,6,7,8-HpCDF	<dl
1,2,3,4,7,8,9-HpCDF	0,022
1,2,3,4,6,7,8,9-OCDF	<0,054
2,3,7,8-TCDD	<0,036
1,2,3,7,8-PeCDD	<0,054
1,2,3,4,7,8-HxCDD	<0,027
1,2,3,6,7,8-HxCDD	<0,108
1,2,3,7,8,9-HxCDD	0,033
1,2,3,4,6,7,8-HpCDD	<0,09
1,2,3,4,6,7,8,9-OCDD	0,28

Table 2: PCB blank

	[pg/column]
PCB-#28	0,42
PCB-#52	0,67
PCB-#101	0,37
PCB-#153	0,33
PCB-#138	<0,261
PCB-#180	0,302
PCB-#81	<0,027
PCB-#77	<0,045
PCB-#126	0,02
PCB-#169	<dl
PCB-#123	0,09
PCB-#118	0,2
PCB-#114	0,047
PCB-#105	<0,081
PCB-#167	0,112
PCB-#156	0,313
PCB-#157	0,16
PCB-#189	0,099

PCDD/F TEQ (2005)	[pg/column]
lower bound	0,06
upper bound	0,08

PCB-TEQ	[pg/column]
lower bound	0,002
upper bound	0,0023
Sum DIN	2,1

QC-Certificate - 19513 - 721718

Table 3: PCDD/F recoveries

	[%]	RSD [%]	
PCDD/F 13C Recoveries [%]	2,3,7,8-TCDF	96	2
	1,2,3,7,8-PeCDF	94	2
	2,3,4,7,8-PeCDF	92	2
	1,2,3,4,7,8-HxCDF	90	5
	1,2,3,6,7,8-HxCDF	99	4
	2,3,4,6,7,8-HxCDF	100	5
	1,2,3,7,8,9-HxCDF	97	5
	1,2,3,4,6,7,8-HpCDF	105	4
	1,2,3,4,7,8,9-HpCDF	100	4
	1,2,3,4,6,7,8,9-OCDF	99	6
	2,3,7,8-TCDD	94	1
	1,2,3,7,8-PeCDD	91	2
	1,2,3,4,7,8-HxCDD	98	4
	1,2,3,6,7,8-HxCDD	85	4
	1,2,3,7,8,9-HxCDD	99	5
	1,2,3,4,6,7,8-HpCDD	98	4
	1,2,3,4,6,7,8,9-OCDD	95	5

Table 4: PCB recoveries

	[%]	RSD [%]	
PCB 13C Recoveries [%]	PCB-#28	99	1
	PCB-#52	99	2
	PCB-#101	98	1
	PCB-#153	92	1
	PCB-#138	95	1
	PCB-#180	92	1
	PCB-#81	87	2
	PCB-#77	89	3
	PCB-#126	86	3
	PCB-#169	81	3
	PCB-#123	90	7
	PCB-#118	84	11
	PCB-#114	91	4
	PCB-#105	86	8
	PCB-#167	74	9
	PCB-#156	84	7
	PCB-#157	79	8
	PCB-#189	78	7