

Quality Control Certificate

Product: **Smart Column**

Product No.: 19513

Lot No.: **721941**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The Smart Column is part of a 3-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:	PCDD/F-TEQ:	0,09	pg/column
		(crit: <	0,70 pg/column)
	dl-PCB-TEQ:	0,0212	pg/column
		(crit: <	0,05 pg/column)
	Sum Total PCB:	3	pg/column
		(crit: <	300 pg/column)

Results Recoveries:	PCDD/F	89	to	106	%	(crit: 70	to	120	%)
	PCB	84	to	100	%	(crit: 70	to	120	%)

This is to certify that the Smart Column, Lot 721941, passed the required test specifications and is released for sale.

date: 12.05.2025

sign.: 

The company LCTech GmbH is certified according to ISO 9001



QC-Certificate - 19513 - 721941

Hazards:	<p>NOT FOR HUMAN OR DRUG USE!</p> <p>The Smart Column is designed and prepared for usage with the Alumina/Florisil Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.</p>
Quality Control:	<p>All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.</p>
Quality Management:	<p>This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)</p>
Documentation / Data Attached:	<p>table 1 & 2: blankvalues of PCDD/F and PCB table 3 & 4: 13C-Recoveries of PCDD/F and PCB</p>
Analytics	<p>This is to certify that the Smart Column, Lot , passed the required test specifications and is released for sale.</p>
Remarks	<p>Our suppliers maintain the highest standard of quality, however due to the high temperature necessary for several steps in the production, some small charred particles may be visible within a batch of silica or filters without any effect on the clean-up.</p>

QC-Certificate - 19513 - 721941

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 6

Table 1: PCDD/F blank

	[pg/column]
2,3,7,8-TCDF	<0,036
1,2,3,7,8-PeCDF	0,05
2,3,4,7,8-PeCDF	<0,081
1,2,3,4,7,8-HxCDF	<0,027
1,2,3,6,7,8-HxCDF	0,02
2,3,4,6,7,8-HxCDF	<0,045
1,2,3,7,8,9-HxCDF	<0,045
1,2,3,4,6,7,8-HpCDF	0,14
1,2,3,4,7,8,9-HpCDF	0,035
1,2,3,4,6,7,8,9-OCDF	<0,054
2,3,7,8-TCDD	<dl
1,2,3,7,8-PeCDD	<0,054
1,2,3,4,7,8-HxCDD	0,051
1,2,3,6,7,8-HxCDD	<0,108
1,2,3,7,8,9-HxCDD	0,028
1,2,3,4,6,7,8-HpCDD	0,16
1,2,3,4,6,7,8,9-OCDD	1,39

Table 2: PCB blank

	[pg/column]
PCB-#28	0,98
PCB-#52	1,11
PCB-#101	0,36
PCB-#153	0,23
PCB-#138	<dl
PCB-#180	0,348
PCB-#81	0,11
PCB-#77	0,19
PCB-#126	0,1634
PCB-#169	0,161
PCB-#123	0,09
PCB-#118	0,18
PCB-#114	0,016
PCB-#105	0,1
PCB-#167	0,204
PCB-#156	0,173
PCB-#157	0,05
PCB-#189	0,125

PCDD/F TEQ (2005)	[pg/column]
lower bound	0,08
upper bound	0,09

PCB-TEQ	[pg/column]
lower bound	0,0212
upper bound	0,0212
Sum DIN	3

QC-Certificate - 19513 - 721941

Table 3: PCDD/F recoveries

		[%]	RSD [%]
PCDD/F 13C Recoveries [%]	2,3,7,8-TCDF	93	4
	1,2,3,7,8-PeCDF	93	2
	2,3,4,7,8-PeCDF	89	3
	1,2,3,4,7,8-HxCDF	97	2
	1,2,3,6,7,8-HxCDF	103	3
	2,3,4,6,7,8-HxCDF	101	4
	1,2,3,7,8,9-HxCDF	105	3
	1,2,3,4,6,7,8-HpCDF	106	3
	1,2,3,4,7,8,9-HpCDF	94	4
	1,2,3,4,6,7,8,9-OCDF	102	5
	2,3,7,8-TCDD	89	3
	1,2,3,7,8-PeCDD	90	4
	1,2,3,4,7,8-HxCDD	103	3
	1,2,3,6,7,8-HxCDD	90	3
	1,2,3,7,8,9-HxCDD	105	4
	1,2,3,4,6,7,8-HpCDD	101	3
	1,2,3,4,6,7,8,9-OCDD	94	5

Table 4: PCB recoveries

		[%]	RSD [%]
PCB 13C Recoveries [%]	PCB-#28	98	6
	PCB-#52	94	4
	PCB-#101	99	5
	PCB-#153	96	9
	PCB-#138	100	0
	PCB-#180	100	4
	PCB-#81	88	3
	PCB-#77	89	3
	PCB-#126	87	6
	PCB-#169	88	6
	PCB-#123	95	9
	PCB-#118	91	12
	PCB-#114	99	7
	PCB-#105	93	7
	PCB-#167	84	13
	PCB-#156	88	8
	PCB-#157	86	10
	PCB-#189	84	9