

Quality Control Certificate

Product: **EVOLUTION Alox Column**
 Product No.: 20087
 Lot No.: **719978**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The EVOLUTION Alumina Column is part of a 3-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and polychlorinated biphenyl (PCB) congeners.

Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:	PCDD/F-TEQ:	0,19	pg/column
		(crit: <	0,7 pg/column)
	dl-PCB-TEQ:	0,0112	pg/column
		(crit: <	0,05 pg/column)
	Sum Total PCB:	12	pg/column
		(crit: <	300 pg/column)

Results Recoveries:	PCDD/F	82	to	101	%	(crit: 70	to	120	%)
	PCB	86	to	108	%	(crit: 70	to	120	%)

This is to certify that the EVOLUTION Alox Column, Lot 719978, passed the required test specifications and is released for sale.

date: 22.03.2024 sign.: *M. Bradis*

The company LCTech GmbH is certified according to ISO 9001



QC-Certificate - 20087 - 719978

Hazards:	<p>NOT FOR HUMAN OR DRUG USE!</p> <p>The Alumina Column is designed and prepared for usage with the Universal/standard & Smart Column and Carbon Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.</p>
Quality Control:	<p>All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.</p>
Quality Management:	<p>This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)</p>
Documentation / Data Attached:	<p>table 1 & 2: blankvalues of PCDD/F and PCB table 3 & 4: 13C-Recoveries of PCDD/F and PCB</p>
Analytics	<p>This is to certify that the EVOLUTION Alumina Column, Lot , passed the required test specifications and is released for sale.</p>
Remarks	<p>n/a</p>



QC-Certificate - 20087 - 719978

Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Blanks: n= 6

Table 1: PCDD/F blank

	[pg/column]
2,3,7,8-TCDF	<0,036
1,2,3,7,8-PeCDF	0,1
2,3,4,7,8-PeCDF	0,11
1,2,3,4,7,8-HxCDF	0,033
1,2,3,6,7,8-HxCDF	<0,018
2,3,4,6,7,8-HxCDF	<0,045
1,2,3,7,8,9-HxCDF	0,05
1,2,3,4,6,7,8-HpCDF	0,08
1,2,3,4,7,8,9-HpCDF	0,05
1,2,3,4,6,7,8,9-OCDF	<0,054
2,3,7,8-TCDD	0,07
1,2,3,7,8-PeCDD	<0,054
1,2,3,4,7,8-HxCDD	0,05
1,2,3,6,7,8-HxCDD	<0,108
1,2,3,7,8,9-HxCDD	<0,027
1,2,3,4,6,7,8-HpCDD	0,12
1,2,3,4,6,7,8,9-OCDD	0,41

Table 2: PCB blank

	[pg/column]
PCB-#28	3,72
PCB-#52	5,88
PCB-#101	1,56
PCB-#153	0,65
PCB-#138	0,22
PCB-#180	<dl
PCB-#81	0,04
PCB-#77	0,233
PCB-#126	0,0751
PCB-#169	0,122
PCB-#123	<dl
PCB-#118	0,02
PCB-#114	<dl
PCB-#105	<dl
PCB-#167	<0,0027
PCB-#156	<0,027
PCB-#157	0,01
PCB-#189	0,004

PCDD/F TEQ (2005)	[pg/column]
lower bound	0,19
upper bound	0,19

PCB-TEQ	[pg/column]
lower bound	0,0112
upper bound	0,0112
Sum DIN	12

Table 3: PCDD/F recoveries

	[%]	RSD [%]	
PCDD/F 13C Recoveries [%]	2,3,7,8-TCDF	85	8
	1,2,3,7,8-PeCDF	87	10
	2,3,4,7,8-PeCDF	84	10
	1,2,3,4,7,8-HxCDF	90	8
	1,2,3,6,7,8-HxCDF	99	8
	2,3,4,6,7,8-HxCDF	97	10
	1,2,3,7,8,9-HxCDF	96	10
	1,2,3,4,6,7,8-HpCDF	99	9
	1,2,3,4,7,8,9-HpCDF	101	5
	1,2,3,4,6,7,8,9-OCDF	97	10
	2,3,7,8-TCDD	82	6
	1,2,3,7,8-PeCDD	87	9
	1,2,3,4,7,8-HxCDD	95	8
	1,2,3,6,7,8-HxCDD	83	8
	1,2,3,7,8,9-HxCDD	100	9
	1,2,3,4,6,7,8-HpCDD	94	10
	1,2,3,4,6,7,8,9-OCDD	93	12

Table 4: PCB recoveries

	[%]	RSD [%]	
PCB 13C Recoveries [%]	PCB-#28	108	7
	PCB-#52	104	13
	PCB-#101	97	5
	PCB-#153	98	2
	PCB-#138	94	5
	PCB-#180	89	4
	PCB-#81	94	0
	PCB-#77	94	0
	PCB-#126	93	0
	PCB-#169	86	0
	PCB-#123	95	2
	PCB-#118	95	4
	PCB-#114	93	4
	PCB-#105	92	4
	PCB-#167	97	9
	PCB-#156	95	5
	PCB-#157	97	3
	PCB-#189	93	4