

Quality Control Certificate

Product: **209PCB**
 Product No.: 20325
 Lot No.: **720554**

Storage Recommendations: Store the column at room temperature below 25°C

Description: The 209PCB Column is part of a 3-column setup used for the sample preparation of environmental-, food- / feed- and similar matrices with DEXTech systems from LCTech for the analysis of polychlorinated dibenzo-p-dioxins (PCDD), polychlorinated dibenzofurans (PCDF) and all 209 polychlorinated biphenyl (PCB) congeners.

Quality Control Release Inspection and Test Specification

Test Procedure: A solvent blank, spiked with quantification standard has been cleaned on a DEXTech Plus system, spiked with recovery standard, evaporated with the D-EVA and has been quantified with a HRGC/HRMS DFS from Thermo Fisher Scientific at a resolution of R > 10000.

Results Blank Value:

PCDD/F-TEQ:	0,01	pg/column				(crit: 0,7 pg/column)
dl-PCB-TEQ:	0,044	pg/column				(crit: 0,05 pg/column)
Sum total PCB:	26,2	pg/column				(crit: 300 pg/column)

Results Recoveries:

PCDD/F	73	to	110	%	(crit: 45	to	130	%)
PCB	57	to	125	%	(crit: 45	to	130	%)

This is to certify that the 209PCB, Lot 20325, passed the required test specifications and is released for sale.

date: 04.07.2024 sign.: 

The company LCTech GmbH is certified according to ISO 9001



Hazards:	<p>NOT FOR HUMAN OR DRUG USE!</p> <p>The 209 Column is designed and prepared for usage with the Alumina Column from LCTech and for laboratory use only. This product should only be used by qualified personnel familiar with its potential hazards and trained in the handling of hazardous chemicals. Due care should be exercised to prevent unnecessary human contact or ingestion, all procedures should be carried out with suitable gloves, eye protection, and clothing should be worn at all times. Waste should be disposed according to national and regional regulations.</p>
Quality Control:	<p>All ingredients are traceable to certified lots of our supplier. In addition, any ingredient with a new lot will be checked on contamination and efficiency before releasing for production. Monitoring the ongoing production, several columns are chosen at random day for analysis to check on contamination and efficiency.</p>
Quality Management:	<p>This product was produced using a Quality Management System registered to the ISO 9001:2015 (DEKRA)</p>
Documentation / Data Attached:	<p>table 1 & 2: blankvalues of PCDD/F and PCB table 3 & 4: 13C-Recoveries of PCDD/F and PCB</p>
Analytics	<p>All the Columns (n>5) have to perform a clean-up of a solvent blank (10 mL n-hexane), spiked with a 13C - labelled quantifier-standard solution with a default alumina plus or pure 209 method onto a DEXTech Pure or Plus system. There are 2 fractions, fraction 1 (all 209 PCB) and fraction 2 (PCDD/F). Both fractions are spiked with the corresponding 13C - labelled recovery- standard solutions and evaporated with the D-EVA vacuum centrifuge. The extracts are measured with a HRMS-DFS from Thermo Fisher Scientific with a resolution of R > 10000. The HRGCs are equipped with 60 m DB5 MS Columns.</p>
Remarks	<p>Our suppliers maintain the highest standard of quality, however due to the high temperature necessary for several steps in the production, some small charred particles may be visible within a batch of Florisil or filters without any effect on the clean-up.</p>



Results:

Lockmass check: No significant disturbances, or indicators for contaminations are detected.

Table 1: PCB recoveries

	[%]	RSD [%]
PCB#1L	91	9
PCB#3L	97	5
PCB#4L	100	13
PCB#8L	101	2
PCB#15L	102	5
PCB#19L	99	7
PCB#28L	125	5
PCB#54L	104	7
PCB#52L	124	4
PCB#70L	91	7
PCB#81L	109	2
PCB#77L	76	2
PCB#104L	87	4
PCB#95LL	98	11
PCB#101L	87	3
PCB#123L	94	4
PCB#118L	95	4
PCB#114L	95	4
PCB#105L	100	6
PCB#126L	87	5
PCB#155L	91	9
PCB#153L	97	7
PCB#138L	100	6
PCB#167L	101	2
PCB#156L	102	4
PCB#157L	99	3
PCB#169L	125	3
PCB#180L	104	4
PCB#170L	124	5
PCB#188L	91	4
PCB#189L	109	3
PCB#202L	76	5
PCB#205L	87	2
PCB#208L	98	4
PCB#209L	87	3

Table 2: PCB blank

	[pg/column]
PCB#1	<dl
PCB#3	<dl
PCB#4	<dl
PCB#8/5	<dl
PCB#15	<dl
PCB#19	<dl
PCB#28	<dl
PCB#54	<dl
PCB#52/69	<dl
PCB#70	<1,794
PCB#81	<dl
PCB#77	<0,67
PCB#104	<dl
PCB#102/93/98/95	<dl
PCB#101	<dl
PCB#123	<dl
PCB#118	<dl
PCB#114	<dl
PCB#105	<dl
PCB#126	<0,577
PCB#155	<dl
PCB#153	<dl
PCB#138	<dl
PCB#167	<0,285
PCB#156	<dl
PCB#157	<dl
PCB#169	<dl
PCB#180	<dl
PCB#170	<dl
PCB#188	<dl
PCB#189	<dl
PCB#202	<dl
PCB#205	<dl
PCB#208	<dl
PCB#209	<dl

Blanks: n = 6

	[pg/column]
PCB-TEQ	
lower bound	0,044
upper bound	0,044
Sum DIN	1,4

	[pg/column]
Grade of chlorination	
sum mono	0,1855
sum di	13,5218
sum tri	7,2397
sum tetra	2,286
sum penta	1,2562
sum hexa	1,0205
sum hepta	0,4222
sum octa	0,2217
sum nona	0,0123
sum deca	0,0172
sum total	26,183

Blanks: n = 6

Table 3: PCDD/F recoveries

	[%]	
PCDD/F 13C Recoveries [%]	2,3,7,8-TCDF	77
	1,2,3,7,8-PeCDF	75
	2,3,4,7,8-PeCDF	77
	1,2,3,4,7,8-HxCDF	88
	1,2,3,6,7,8-HxCDF	93
	2,3,4,6,7,8-HxCDF	103
	1,2,3,7,8,9-HxCDF	108
	1,2,3,4,6,7,8-HpCDF	104
	1,2,3,4,7,8,9-HpCDF	87
	1,2,3,4,6,7,8,9-OCDF	96
	2,3,7,8-TCDD	73
	1,2,3,7,8-PeCDD	82
	1,2,3,4,7,8-HxCDD	109
	1,2,3,6,7,8-HxCDD	93
	1,2,3,7,8,9-HxCDD	110
	1,2,3,4,6,7,8-HpCDD	94
	1,2,3,4,6,7,8,9-OCDD	87

Table 4: PCDD/F blank

	[pg/column]	
native amount	2,3,7,8-TCDF	<dl
	1,2,3,7,8-PeCDF	<0,045
	2,3,4,7,8-PeCDF	<dl
	1,2,3,4,7,8-HxCDF	0,041
	1,2,3,6,7,8-HxCDF	0,029
	2,3,4,6,7,8-HxCDF	<0,045
	1,2,3,7,8,9-HxCDF	<0,045
	1,2,3,4,6,7,8-HpCDF	0,26
	1,2,3,4,7,8,9-HpCDF	<dl
	1,2,3,4,6,7,8,9-OCDF	<0,054
	2,3,7,8-TCDD	<dl
	1,2,3,7,8-PeCDD	<dl
	1,2,3,4,7,8-HxCDD	<dl
	1,2,3,6,7,8-HxCDD	<dl
	1,2,3,7,8,9-HxCDD	<dl
	1,2,3,4,6,7,8-HpCDD	<dl
	1,2,3,4,6,7,8,9-OCDD	0,4

PCDD/F TEQ (2005)	[pg/column]
lower bound	0,01
upper bound	0,06